

---

## **TWL Calibration Data Summary**

---

**Serial Number:** 0000959

**Full-Scale Flow Rate:** 80  
(yields 20mA output)

**Zero Flow Rate:** (yields 4mA output)

**Bridge Resistance:** 351.1 ohms

**Full-Scale Bridge Output:** 1.9414 mV/V

**Calibration Resistor (R7):** 47297 ohms

**Calibration Voltage:** 0.9524 volts

**Calibration Flow Rate:** 78.07

**Calibration Current:** 19.61 mA

**Zero Switch Setting:** 0

**TB3 Jumper Position:** + (POSITIVE)

**Span Switch Setting - Undamped:** 0

**Span Switch Setting - Damped:** 8

**Additional Information:**

---